



**SCREEN IFA TEST MIOGLOBINA
(Whole Blood /Serum/Plasma)
Package Insert**

REF: SC-1484	English
--------------	---------

A rapid test for the diagnosis of myocardial infarction (MI) by measuring Myoglobin in whole blood, Serum or plasma with the use of SCREEN® fluorescence Immunoassay Analyzer. For professional in vitro diagnostic use only.

INTENDED USE

The Myoglobin Test Cassette (Whole Blood/Serum/Plasma) is intended for in vitro quantitative determination of Myoglobin in whole blood, serum or plasma as an aid in the diagnosis of Myocardial Infarction (MI).

SUMMARY

Myoglobin (MYO) is a heme-protein normally found in skeletal and cardiac muscle with a molecular weight of 17.8 kDa. It constitutes about 2 percent of total muscle protein and is responsible for transporting oxygen within the muscle cells.1 When the muscle cells are damaged, Myoglobin is released to the blood rapidly due to its relatively small size. Following the death of tissue associated with MI, Myoglobin is one of the first markers to rise above normal levels. The level of Myoglobin increases measurably above baseline within 2-4 hours post-infarct, peaking at 9-12 hours and returning to baseline within 24-36 hours.2,3 A number of reports suggest the measurement of Myoglobin as a diagnostic aid in confirming the absence of myocardial infarction with negative predictive values of up to 100% reported at certain time periods after onset of symptoms.4

PRINCIPLE

The Myoglobin Test Cassette (Whole Blood/Serum/Plasma) detects Myoglobin based on Fluorescence Immunoassay. The sample moves through the strip from sample pad to absorbent pad. If the specimen contains Myoglobin, it attaches to the fluorescent microspheres-conjugated anti-Myoglobin antibodies. Then the complex will be captured by the capture antibodies coated on the nitrocellulose membrane (Test line). The concentration of Myoglobin in the sample correlates with the fluorescence signal intensity captured on the T line. According to the fluorescence intensity of the test and the standard curve, the concentration of Myoglobin in the sample can be calculated by SCREEN® Analyzer to show Myoglobin concentration in specimen.

REAGENTS

The test includes anti-Myoglobin antibody coated fluorophores and anti-Myoglobin antibody coated on the membrane.

PRECAUTIONS

- For professional *in vitro* diagnostic use only.
- Do not use after the expiration date indicated on the package. Do not use the test if the foil pouch is damaged. Do not reuse.
- Avoid cross-contamination of specimens by using a new specimen collection container for each specimen obtained.
- Do not eat, drink or smoke in the area where the specimens and tests are handled. Handle all specimens as if they contain infectious agents. Observe established precautions against microbiological hazards throughout the procedure and follow standard procedures for proper disposal of specimens. Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are assayed.
- Do not interchange or mix reagents from different lots.
- Humidity and temperature can adversely affect results.
- Used testing materials should be discarded in accordance with local regulations.
- Read the entire procedure carefully prior to any testing.
- The Myoglobin Test Cassette should only be used with the SCREEN® Analyzer by approved medical professionals.

STORAGE AND STABILITY

- The kit should be stored at 4-30 °C until the expiry date printed on the sealed pouch.
- The test must remain in the sealed pouch until use.
- Do not freeze.
- Care should be taken to protect the components of the kit from contamination. Do not use if there is evidence of microbial contamination or precipitation. Biological contamination of dispensing equipment, containers or reagents can lead to false results.

SPECIMEN COLLECTION AND PREPARATION

Preparation

Before performing the test, please make sure that all components are brought to room temperature (15-30 °C). Cold buffer solution or moisture condensation on the membrane can lead to invalid test results.

Sample Handling

- Collect the specimen according to standard procedures.
- Do not leave specimens at room temperature for prolonged periods. Serum/Plasma specimens may be stored at 2-8 °C for up to 3 days, for long term storage, specimens should be kept below -20 °C. Whole blood collected by venipuncture should be stored at 2-8 °C if the test is to be run within 2 days of collection. Do not freeze whole blood specimens. Whole blood collected by finger stick should be tested immediately.
- Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to testing. Avoid repeated freezing and thawing of specimens.
- EDTA K2, Heparin sodium, Citrate sodium and Potassium Oxalate can be used as the anticoagulant for collecting the specimen.

Sample Dilution / Sample Stability

- The specimen (**50 ul of serum / plasma / 75 ul of whole blood**) can be added directly with the micro pipette into the buffer.
- Close the tube and shake the sample by hand vigorously for approximately **10 seconds** to mix the sample and dilution buffer.
- Let the diluted sample homogenize for approximately **1 minute**.
- It is best to place the diluted sample on an ice pack and leave the sample at room temperature for no more than 8 hours.

MATERIALS

Materials Provided

- Test Cassettes
- Specimen Collection Tubes with Extraction Buffer
- ID Card
- Package Insert

Materials Required But Not Provided

- Timer
- Centrifuge
- SCREEN® Fluorescence Immunoassay Analyzer
- Pipette
- Specimen Collection Containers

DIRECTIONS FOR USE

Refer to SCREEN® Fluorescence Immunoassay Analyzer Operation Manual for the complete instructions on use of the Test. The test should be conducted in room temperature. **Allow the test, specimen, buffer and/or controls to reach room temperature (15-30°C) prior to testing.**

- Turn on the Analyzer power. Then according to the need, select "standard test" or "Quick test" mode.
- Take out the ID card and insert it into the Analyzer ID Card Slot.
- Serum/Plasma:** Pipette **50µl serum/plasma** into the buffer tube, mix the specimen and the buffer well.
- Whole blood:** Transfer **75µl whole blood** into the buffer tube with pipette; mix the specimen and the buffer well.
- Add diluted specimen with a Pipette:** Pipette **85µl** diluted specimen into the sample well of the test cassette. Start the timer at the same time.
- There are two test modes for SCREEN® Fluorescence Immunoassay Analyzer, Standard Test mode and Quick Test mode. Please refer to the user manual of SCREEN® Fluorescence Immunoassay Analyzer for details.
 - "Quick test"** mode: After **15 minutes** of adding sample, Insert the test cassette into the Analyzer, click **"QUICK TEST"**, fill the test information and click **"NEW TEST"** immediately. The Analyzer will automatically give the test result after a few seconds.
 - "Standard test"** mode: Insert the test cassette into the Analyzer immediately after adding specimen, click **"STANDARD TEST"**, fill the test information and click **"NEW TEST"** at the

same time, The Analyzer will automatically countdown **15 minutes**. After the countdown, the Analyzer will give the result at once.

INTERPRETATION OF RESULTS

Results read by SCREEN® Fluorescence Immunoassay Analyzer.

The result of tests for Myoglobin is calculated by SCREEN® Fluorescence Immunoassay Analyzer and display the result on the screen. For additional information, please refer to the user manual of SCREEN® Fluorescence Immunoassay Analyze.

Linearity range of Myoglobin Test is 5–200 ng/mL.

QUALITY CONTROL

Each Myoglobin Test Cassette contains internal control that satisfies routine quality control requirements. This internal control is performed each time a patient sample is tested. This control indicates that the test device was inserted and read properly by SCREEN® Fluorescence Immunoassay Analyzer. An invalid result from the internal control causes an error message on SCREEN® Fluorescence Immunoassay Analyzer indicating that the test should be repeated. An invalid result from the internal control causes an "N/A" message on SCREEN® Fluorescence Immunoassay Analyzer. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control failure. Review the procedure and repeat the test with a new test. If the problem persists, discontinue using the test kit immediately and contact your local distributor.

LIMITATIONS

- The Myoglobin Test Cassette (Whole Blood/Serum/Plasma) is for professional in vitro diagnostic use, and should only be used for the quantitative detection of Myoglobin.
- The Myoglobin Test Cassette (Whole Blood/Serum/Plasma) will only indicate the presence of Myoglobin in the specimen and should not be used as the sole criterion for evaluating AMI.
- Like with all diagnostic tests, a confirmed diagnosis should only be made by a physician after all clinical and laboratory findings have been evaluated.
- High concentrations of Myoglobin may produce a dose hook effect, resulting in incorrect interpretation of Myoglobin levels. High dose hook effect has not been observed with this test up to 200 ng/mL of Myoglobin.
- The hematocrit level of the whole blood should be between 25% and 65%.
- The results of Myoglobin Rapid Tests are based on measuring the levels of Myoglobin in a specimen. It should not be used as the sole criterion for treatment decisions. If the result is positive, other clinical findings and alternative test methods are recommended to reach proper medical treatments.

EXPECTED VALUES

Concentrations	Clinical Reference
< 90 ng/mL	Not indicative of Acute Myocardial Infarction
> 90 ng/mL	Indicative of Acute Myocardial Infarction

PERFORMANCE CHARACTERISTICS

1. Accuracy

The test deviation is $\leq \pm 15\%$.

2. Sensitivity

The Myoglobin Test Cassette (Whole Blood/Serum/Plasma) can detect levels of Myoglobin as low as 5 ng/mL in whole blood, serum or plasma.

3. Detection range

5–200 ng/mL

4. Linearity range

5–200 ng/mL , R \geq 0.990

5. Precision

Intra-lot precision

Within-run precision has been determined by using 10 replicates of 5 specimens containing 0 ng/mL, 50 ng/mL, 100 ng/mL and 200 ng/mL of Myoglobin. C.V. is $\leq 15\%$.

Inter-lot precision

Between-run precision has been determined by using 10 replicates for each of three lots using 5 specimens containing 0 ng/mL, 50 ng/mL, 100 ng/mL and 200 ng/mL of Myoglobin. C.V. is $\leq 15\%$.

6. Cross-reactivity

Cross-reactivity studies were carried out with following analytes.
 HBsAg, HBsAb, HBeAg, HBeAb, HbCAb, syphilis, anti-HIV, anti-H.pylori, MONO, anti-CMV,
 anti-Rubella and anti-Toxoplasmosis positive specimens. The results showed no cross-
 reactivity.

7. Interfering Substances

The following potentially interfering substances were added to Myoglobin negative and positive specimens, respectively.

- Acetaminophen:20 mg/dL Caffeine:20 mg/dL
- Acetylsalicylic Acid:20 mg/dL Gentisic Acid:20 mg/dL
- Ascorbic Acid:20mg/dLAlbumin:10,500mg/dL
- Creatin:200 mg/dLHemoglobin:1,000 mg/dL
- Bilirubin:1,000mg/dLOxalic Acid:600mg/dL
- Cholesterol:800mg/dLTriglycerides:1,600mg/dL

None of the substances at the concentration tested interfered in the assay

8. Method comparison

The SCREEN® Myoglobin Test Cassette was compared with the results obtained with Cobas E602 for 98 samples. The correlation coefficient(r) is 0.981.

BIBLIOGRAPHY

- 1.Wong SS. Strategic utilization of cardiac markers for diagnosis of acute myocardial infarction. Ann Clin Lab Sci, 26:301-12, 1996.
- 2.Kagen LJ.Myoglobin methods and diagnostic uses.CRC Crit.Rev. Clin.Lab.Sci., 2:273,1978.
- 3.Chapelle JP.et al.Serum Myoglobin determinations in the assessment of acute myocardial infarction. Eur. Heart Journal, 3:122, 1982.
- 4.Hamfelt A. et al.Use of biochemical tests for myocardial infarction in the county of Vasternorrland, a clinical chemistry routine for the diagnosis of myocardial infarction. Scand. J. Clin. Lab. Invest. Suppl., 200:20, 1990.

INDEX OF SYMBOLS

	Consult Instructions for Use		Tests per kit		Authorized Representative
	For in vitro diagnostic use only		Use by		Do not reuse
	Store between 4-30°C		Lot Number		Catalog #
	Do not use if package is damaged		Manufacturer		

 **SCREEN ITALIA S.r.l.**
 Via dell'Artigianato, 16
 06089 - Torgiano - Perugia - Italia
www.screenitalia.it info@screenitalia.it



Number: 145914201

Effective Date: 25-08-2020